

RBC Chamber Blockage or Constant Flags

Press 4 - Service

Press 2 - Drain Chambers

3mL Distilled Water in Mixing Chamber

3mL 50/50 Distilled water/Bleach in RBC Chamber

Press 5 – Transfer

Press 1 – Back Flush 10 times

Press 2 – Drain Chambers

Escape back to Main Menu

Press the Start Up Button

Run 3 Blood Controls